

MDCG 2023-6

Guidance on demonstration of equivalence for Annex XVI products

A guide for manufacturers and notified bodies

December 2023

This document has been endorsed by the Medical Device Coordination Group (MDCG) established by Article 103 of Regulation (EU) 2017/745. The MDCG is composed of representatives of all Member States and it is chaired by a representative of the European Commission.

The document is not a European Commission document and it cannot be regarded as reflecting the official position of the European Commission. Any views expressed in this document are not legally binding and only the Court of Justice of the European Union can give binding interpretations of Union law.

1 Introduction

The Regulation (EU) 2017/745 on medical devices¹, hereafter referred to as the MDR (medical device regulation), provides a possibility to use clinical data related to an equivalent device in the clinical evaluation required for a device under conformity assessment².

Whilst carrying out a clinical investigation is the most direct way to generate clinical data concerning the safety and performance of devices for the purpose of CE marking, clinical data can also be sourced from data of a device for which the equivalence to the device in question can be demonstrated³. In such cases, equivalence shall be demonstrated according to the MDR requirements⁴.

According to Commission Implementing Regulation (EU) 2022/2346 laying down common specifications for the groups of products without an intended medical purpose listed in Annex XVI to Regulation (EU) 2017/745 on medical devices⁵, hereinafter referred to as CS, in general it is not possible to demonstrate equivalence between a medical device and a product without an intended medical purpose where all available results of clinical investigations relate to medical devices only. Therefore, clinical investigations should be performed for products without an intended medical purpose⁶.

The guidance document MDCG 2020-5 on “Clinical Evaluation – Equivalence”⁷ also covers the products without an intended medical purpose listed in the Annex XVI of MDR⁸. It was issued in April 2020, before the availability of the CS and with the purposes of highlighting the differences between the MDR and the MEDDEV 2.7/1 rev.4⁹.

2 Scope

This MDCG guidance covers the demonstration of equivalence, based on data pertaining to an already existing device, for the purpose of CE-marking under the MDR and is applicable to products without an intended medical purpose listed in the Annex XVI of MDR and covered by the CS. For

¹ Regulation (EU) 2017/745 of the European Parliament and of the Council of 5 April 2017 on medical devices - <http://data.europa.eu/eli/reg/2017/745/2020-04-24>.

² MDR, Article 61 and Annex XIV Part A.

³ MDR, Article 2 (48) 2nd and 3rd indent.

⁴ MDR, Annex XIV, Part A (3).

⁵ Commission Implementing Regulation (EU) 2022/2346 laying down common specifications for the groups of products without an intended medical purpose listed in Annex XVI to Regulation (EU) 2017/745 of the European Parliament and of the Council on medical devices – http://data.europa.eu/eli/reg_impl/2022/2346/oj.

⁶ CS, recital (11).

⁷ https://ec.europa.eu/health/medical-devices-sector/new-regulations/guidance-mdcg-endorsed-documents-and-other-guidance_en#sec2.

⁸ MDCG 2020-5 “Clinical Evaluation – Equivalence”, points (a) and (f) of paragraph 4 on “Demonstration of equivalence”.

⁹ MEDDEV 2.7/1 rev.4 “Clinical evaluation: Guide for manufacturers and notified bodies” in support of the application of Council Directives 93/42/EEC and 90/385/EEC on medical devices and active implantable medical devices, which have been repealed by the MDR.

dual-purpose devices, which are devices with a medical and a non-medical intended purpose, this guidance applies only to the non-medical intended purpose.

This guidance document should be used in conjunction with MDCG 2020-5 on equivalence.

3 Demonstration of equivalence

3.1 Product without an intended medical purpose vs product without an intended medical purpose

When referring to clinical data from an equivalent device, manufacturers should consider that equivalence between two devices without an intended medical purpose should be demonstrated in accordance with the criteria established in the MDR¹⁰. In particular, the demonstration has to consider technical, biological and clinical characteristics.

The technical and biological criteria listed in the MDR can be directly considered to compare characteristics of products without an intended medical purpose. The criterion for the clinical characteristics should be considered taking into account that some of them are specifically referring to a medical purpose. The following table clarifies how the characteristics should be considered for products without an intended medical purpose.

Elements of the Clinical criterion (MDR, Annex XIV Part A Section 3)	Clarifications for the application of the Clinical criterion to products without a medical purpose
the device is used for the same clinical condition or purpose ,	only 'used for the same purpose' applies
including similar severity and stage of disease ,	'similar severity and stage of disease' is not applicable
at the same site in the body,	'at the same site in the body' is directly applicable
in a similar population, including as regards age, anatomy and physiology;	'in a similar population' is directly applicable as well as the similar 'age', 'anatomy' and 'physiology'
has the same kind of user;	'the same kind of user' is directly applicable
has similar relevant critical performance	'similar relevant critical performance' is directly applicable
in view of the expected clinical effect for a specific intended purpose	the expected 'clinical' effect is not applicable. Nevertheless, the products have expected effects for the specific intended purpose, therefore the requirement should be considered as 'in view of the expected effect for a specific intended purpose'

3.2 Product without an intended medical purpose vs analogous medical device

In general, a comparison is not possible between a medical device and a product without an intended medical purpose¹¹ because not all the clinical characteristics can be compared. In particular, the

¹⁰ MDR, Annex XIV Part A (3).

¹¹ CS, recital (11).

characteristic ‘similar severity and stage of disease’ would be defined and available for the medical device while it would not be defined and available for the product without an intended medical purpose¹². As a consequence, the demonstration of equivalence cannot be completed and established between a device without an intended medical purpose and an analogous device with a medical purpose¹³.

3.3 Product without an intended medical purpose vs dual-purpose device

Equivalence between a product without an intended medical purpose and a device with both a medical and a non-medical purpose (dual-purpose device) can be demonstrated comparing the characteristics related to the non-medical purpose for both the devices. For the dual-purpose device, only the characteristics related to the non-medical purpose should be considered.

If equivalence is demonstrated, only clinical data of the dual-purpose device related to the general safety and performance requirements¹⁴ applicable for the non-medical purpose should be used for the clinical evaluation of the product without an intended medical purpose.

¹² This further explains the sentence “To duly justify reliance on existing clinical data from an analogous medical device, the principles of demonstration of equivalence should be applied with the acceptance that the device under evaluation will only have an aesthetic or another non-medical purpose whereas the analogous device has a medical purpose” from point 4(f) of MDCG 2020-5.

¹³ In cases where equivalence cannot be demonstrated, data from similar devices may be useful for a variety of other purposes (e.g., ensuring that the risk management system is comprehensive; understanding the state of the art; helping to define the scope of the clinical evaluation) that are listed in point 5 of MDCG 2020-5.

¹⁴ MDR, Annex I.